

Contractor:

G T Michelli
130 Brookhollow Esplanade
Harahan, LA 70123 5102

Purchase Order #:

11016440

Client:

G T Michelli

Address:

130 Brookhollow Esplanade

City & State:

Harahan, LA 70123 5102

Date Received:

07 OCT 2008

Date Calibrated:

08 OCT 2008 to 09 OCT 2008

Temperature Range:

21.92 to 22.75 °C

Pressure Range:

725.2 to 729.6 mmHg

Relative Humidity Range:

49 to 50 %

Air Density:

1.1333 to 1.1426 mg/cm³

Traceable Report #:

1459884

NIST Certificate #:

822/272801-06 822/274081-06

Tested By:

12

Procedure:

Modified Substitution (WI05-0023)

Contractor Req Recall Date:

2 Year

Primary Standard Calibration Date:

11/03/06

Due: 11/03/10

Description of Weights:

5mg-100g Polished Kit, Class "S", S/N 1228784

Although there are two NIST numbers, one or both may apply.



Indicates As Found >= Tol

Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
5 mg		-0.00099	-0.00099	0.00067	0.014	501Q	K594Q 09-05-08/12-05-08	2.70
10 mg		0.0086	0.0086	0.0010	0.014	501Q	K594Q 09-05-08/12-05-08	2.70
20 mg		0.00172	0.00172	0.00081	0.014	501Q	K594Q 09-05-08/12-05-08	2.70
20 mg		-0.00400	-0.00400	0.00081	0.014	501Q	K594Q 09-05-08/12-05-08	2.70
50 mg		0.00192	0.00192	0.00100	0.014	501Q	K594Q 09-05-08/12-05-08	7.95
50 mg		0.00506	0.00506	0.00100	0.014	501Q	K594Q 09-05-08/12-05-08	7.95
100 mg		-0.0185	-0.0185	0.0014	0.025	501Q	K594Q 09-05-08/12-05-08	7.95
200 mg		0.0139	0.0139	0.0013	0.025	501Q	K594Q 09-05-08/12-05-08	7.95
200 mg		-0.0142	-0.0142	0.0013	0.025	501Q	K594Q 09-05-08/12-05-08	8.00
500 mg		-0.0049	-0.0049	0.0016	0.025	501Q	K594Q 09-05-08/12-05-08	7.95
1 g		0.0159	0.0159	0.0023	0.054	501Q	K594Q 09-05-08/12-05-08	8.00
2 g		0.1062	0.0176	<input checked="" type="checkbox"/> 0.0026	0.054	501Q	K594Q 09-05-08/12-05-08	8.00
2 g		0.0358	0.0358	0.0026	0.054	501Q	K594Q 09-05-08/12-05-08	8.00
5 g		0.0928	0.0127	<input checked="" type="checkbox"/> 0.0055	0.054	501Q	K594Q 09-05-08/12-05-08	8.00
10 g		-0.023	-0.023	0.012	0.074	676Q	K594Q 09-05-08/12-05-08	8.00
20 g		0.043	0.043	0.012	0.074	676Q	K594Q 09-05-08/12-05-08	8.00
20 g		0.000	0.000	0.012	0.074	676Q	K594Q 09-05-08/12-05-08	8.00
50 g		0.039	0.039	0.020	0.12	1183Q	K594Q 09-05-08/12-05-08	8.00
100 g		-0.068	-0.068	0.043	0.25	1183Q	K594Q 09-05-08/12-05-08	8.00

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:



Dated 09 OCT 2008

Debbie Schieffer
Debbie Schieffer Lab Supervisor

230 West Coleman Street • Rice Lake, WI 54868 • USA
TEL: 715-234-9171 • FAX: 715-234-6967

This report is not to be used to claim product endorsement by Rice Lake Weighing Systems or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory. The above uncertainty does not include a component for magnetic property or handling and use.