

**Contractor:**

**G T Michelli Co Inc**  
130 Brookhollow Esplanade  
Harahan, LA 70123 5102

**Purchase Order #:**

11016773

**Client:**

**G T Michelli Co Inc**

**Address:**

130 Brookhollow Esplanade

**City & State:**

Harahan, LA 70123 5102

**Date Received:**

23 FEB 2009

**Date Calibrated:**

25 FEB 2009 to 26 FEB 2009

**Temperature Range:**

21.96 to 22.21 °C

**Pressure Range:**

729.7 to 733.2 mmHg

**Relative Humidity Range:**

49 to 50 %

**Air Density:**

1.1421 to 1.1473 mg/cm³

**Traceable Report #:**

1495822

**NIST Certificate #:**

822/272801-06 822/274081-06

**Tested By:**

17

**Procedure:**

Modified Substitution (WI05-0023)

**Contractor Req Recall Date:**

2 Year

**Primary Standard Calibration Date:**

11/03/06

**Due:** 11/03/10

**Description of Weights:**

1g-1kg Polished Kit, Class "S", S/N 1228785

Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used	Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm³)
		As Found (mg)	As Left (mg)						
1 g		0.0180	0.0180	0.0023	0.054	501Q	K594Q	12-02-08/03-02-09	7.85
2 g		0.0282	0.0282	0.0027	0.054	501Q	K594Q	12-02-08/03-02-09	7.85
2 g		0.0280	0.0280	0.0027	0.054	501Q	K594Q	12-02-08/03-02-09	7.85
5 g		0.0414	0.0414	0.0057	0.054	501Q	K594Q	12-02-08/03-02-09	7.85
10 g		-0.054	-0.054	0.011	0.074	676Q	K594Q	12-02-08/03-02-09	7.85
20 g		0.005	0.005	0.011	0.074	676Q	K594Q	12-02-08/03-02-09	7.85
20 g		-0.056	-0.056	0.011	0.074	676Q	K594Q	12-02-08/03-02-09	7.85
50 g		0.093	0.093	0.019	0.12	1183Q	K594Q	12-02-08/03-02-09	7.85
100 g		-0.028	-0.028	0.043	0.25	1183Q	K594Q	12-02-08/03-02-09	7.85
200 g		0.133	0.133	0.054	0.50	619Q	K594Q	12-02-08/03-02-09	7.85
200 g		0.056	0.056	0.054	0.50	619Q	K594Q	12-02-08/03-02-09	7.85
500 g		0.996	0.996	0.076	1.2	619Q	K594Q	12-02-08/03-02-09	7.85
1 kg		2.27	2.27	0.15	2.5	619Q	K594Q	12-02-08/03-02-09	7.85

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:



230 West Coleman Street • Rice Lake, WI 54868 • USA  
TEL: 715-234-9171 • FAX: 715-234-6967

Page 1 of 1 Page

Dated 26 FEB 2009

*Dan Demers*  
Dan Demers Metrologist

This report is not to be used to claim product endorsement by Rice Lake Weighing Systems or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory. The above uncertainty does not include a component for magnetic property or handling and use.