

**Contractor:**

**G T Michelli**  
130 Brookhollow Esplanade  
Harahan, LA 70123 5102

**Purchase Order #:**

11016875

**Client:**

**G T Michelli**

**Address:**

130 Brookhollow Esplanade

**City & State:**

Harahan, LA 70123 5102

**Date Received:**

08 APR 2009

**Date Calibrated:**

10 APR 2009

**Temperature Range:**

21.81 to 22.20 °C

**Pressure Range:**

735.0 to 735.1 mmHg

**Relative Humidity Range:**

49 to 50 %

**Air Density:**

1.1500 to 1.1516 mg/cm<sup>3</sup>

**Traceable Report #:**

1508494

**NIST Certificate #:**

822/272801-06 822/274081-06

**Tested By:**

12

**Procedure:**

Modified Substitution (WI05-0023)

**Contractor Req Recall Date:**

2 Year

**Primary Standard Calibration Date:**

11/03/06

**Due:** 11/03/10

**Description of Weights:**

100g-5kg Satin Finish Weights, Class "2", S/N U084

Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm <sup>3</sup> )
		As Found (mg)	As Left (mg)						
100 g		0.008	0.008	0.040	0.50	1183Q	K594Q	03-02-09/06-02-09	7.84
200 g	1	-0.117	-0.117	0.067	1.0	619Q	K594Q	03-02-09/06-02-09	7.84
200 g	2	0.328	0.328	0.067	1.0	619Q	K594Q	03-02-09/06-02-09	7.84
500 g	1	-0.763	-0.763	0.080	2.5	619Q	K594Q	03-02-09/06-02-09	7.84
500 g	2	-0.788	-0.788	0.080	2.5	619Q	K594Q	03-02-09/06-02-09	7.84
1 kg	1	-1.82	-1.82	0.15	5.0	619Q	K594Q	03-02-09/06-02-09	7.84
1 kg	2	-1.61	-1.61	0.15	5.0	619Q	K594Q	03-02-09/06-02-09	7.84
2 kg	1	-4.32	-4.32	0.52	10	975Q	K594Q	03-02-09/06-02-09	7.84
5 kg	1	-5.16	-5.16	0.89	25	975Q	K594Q	03-02-09/06-02-09	7.84
5 kg	2	6.44	6.44	0.89	25	975Q	K594Q	03-02-09/06-02-09	7.84

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:



230 West Coleman Street • Rice Lake, WI 54868 • USA  
TEL: 715-234-9171 • FAX: 715-234-6967

Dated 10 APR 2009

*Dan Demers*  
Dan Demers Metrologist

This report is not to be used to claim product endorsement by Rice Lake Weighing Systems or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory. The above uncertainty does not include a component for magnetic property or handling and use.