

Contractor: G T Michelli Co Inc
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 41685
Client: G T Michelli Co
Address: 9663 Mammoth Dr

City & State: Baton Rouge, LA 70814
Date Received: 11 AUG 2009
Date Calibrated: 14 AUG 2009
Temperature Range: 21.28 to 21.75 °C
Pressure Range: 731.2 to 731.3 mmHg
Relative Humidity Range: 50 %
Air Density: 1.1460 to 1.1479 mg/cm³
Traceable Report #: 1540993A
NIST Certificate #: 822/272801-06 822/274081-06
Tested By: 21
Procedure: Mass Dissemination (reference HB 952)
Contractor Req Recall Date: 2 Year
Primary Standard Calibration Date: 11/03/06 **Due:** 11/03/10
Description of Weights: 6-25kg Polished Grip Handle Wt. Class 1



Although there are two NIST numbers, one or both may apply.

Nominal Value	Id.	Conventional Mass Corr.		Unc. ¹ K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
25 kg	2JPG	26.7	26.7	9.8	62	502Q	L595Q 08-10-09/11-10-09	7.84
25 kg	2JPH	15.5	15.5	9.8	62	502Q	L595Q 08-10-09/11-10-09	7.84
25 kg	2JPI	18.7	18.7	9.6	62	502Q	L595Q 08-10-09/11-10-09	7.84
25 kg	2JPJ	17.5	17.5	9.6	62	502Q	L595Q 08-10-09/11-10-09	7.84
25 kg	2NSV	20.4	20.4	9.6	62	502Q	L595Q 08-10-09/11-10-09	7.84
25 kg	2NSW	22.7	22.7	9.6	62	502Q	L595Q 08-10-09/11-10-09	7.84

¹ Uncertainties apply to As Left values only

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.



230 West Coleman Street
 Rice Lake, WI 54668 • USA
 TEL: 715-234-9171 • FAX: 715-234-8967

Dated 18 AUG 2009

Dan Demers
 Dan Demers Metrologist

