

Contractor: G T Michelli Co Inc
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 41685
Client: G T Michelli Co
Address: 9663 Mammoth Dr

City & State: Baton Rouge, LA 70814
Date Received: 11 AUG 2009
Date Calibrated: 14 AUG 2009
Temperature Range: 21.78 to 22.28 °C
Pressure Range: 730.2 to 730.3 mmHg
Relative Humidity Range: 50 to 55 %
Air Density: 1.1416 to 1.1439 mg/cm³
Traceable Report #: 1540993B
NIST Certificate #: 822/272601-06 822/274081-06
Tested By: 21
Procedure: Mass Dissemination (reference HB 952)
Contractor Req Recall Date: 2 Year
Primary Standard Calibration Date: 11/03/06 **Due:** 11/03/10
Description of Weights: 10g-5kg Satin Finish Weights, Class "S1", S/N 2B5W

Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. ¹ K=2 (mg)	Tot. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
10 g	.	-0.052	-0.062	0.011	0.25	1180Q	L595Q 08-10-09/11-10-09	7.84
50 g	0YED	0.026	0.026	0.021	0.60	1470Q	L595Q 08-10-09/11-10-09	7.84
100 g	0YEE	0.207	0.207	0.045	1.0	1470Q	L595Q 08-10-09/11-10-09	7.84
200 g	2B50	0.369	0.369	0.052	2.0	699Q	L595Q 08-10-09/11-10-09	7.84
1 kg	0YEG	-6.81	-6.81	0.11	10	699Q	L595Q 08-10-09/11-10-09	7.84
2 kg	0YEH	6.66	6.66	0.32	20	124Q	L595Q 08-10-09/11-10-09	7.84
2 kg	0YEI	-9.91	-9.91	0.32	20	124Q	L595Q 08-10-09/11-10-09	7.84
5 kg	MY	18.56	18.56	0.80	50	124Q	L595Q 08-10-09/11-10-09	7.84

¹ Uncertainties apply to As Left values only

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.



230 West Coleman Street
 Rice Lake, WI 54956 • USA
 TEL: 715-234-9171 • FAX: 715-234-6967

Dated 14 AUG 2009

Dan Demers
 Dan Demers Metrologist

