

Contractor: G T Michelli Co Inc
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 41685
Client: G T Michelli Co
Address: 9663 Mammoth Dr

City & State: Baton Rouge, LA 70814
Date Received: 11 AUG 2009
Date Calibrated: 14 AUG 2009
Temperature Range: 21.80 to 21.84 °C
Pressure Range: 730.3 to 730.4 mmHg
Relative Humidity Range: 51 %
Air Density: 1.1440 to 1.1442 mg/cm³
Traceable Report #: 1540993D
NIST Certificate #: 822/272801-06 822/274081-06
Tested By: 21
Procedure: Mass Dissemination (reference HB 952)
Contractor Req Recall Date: 2 Year
Primary Standard Calibration Date: 11/03/06 **Due:** 11/03/10
Description of Weights: 2-1kg, 2kg Polished Wts, Class 1
 Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. ¹ K=2 (mg)	Tot. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
1 kg	2JPE	0.213	0.213	0.098	2.5	699Q	L595Q 08-10-09/11-10-09	7.84
1 kg	2JPF	0.389	0.389	0.098	2.5	699Q	L595Q 08-10-09/11-10-09	7.84
2 kg	2JPA	1.27	1.27	0.42	5.0	124Q	L595Q 08-10-09/11-10-09	7.84

¹ Uncertainties apply to As Left values only

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.



230 West Coleman Street
 Rice Lake, WI 54868 • USA
 TEL: 715-234-9171 • FAX: 715-234-6067

Dated 14 AUG 2009

Dan Demers
 Dan Demers Metrologist

