

Contractor: G T Michelli Co Inc
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 41685
Client: G T Michelli Co
Address: 9663 Mammoth Dr

City & State: Baton Rouge, LA 70814
Date Received: 11 AUG 2009
Date Calibrated: 14 AUG 2009 to 17 AUG 2009
Temperature Range: 21.17 to 22.20 °C
Pressure Range: 730.3 to 730.6 mmHg
Relative Humidity Range: 51 to 53 %
Air Density: 1.1424 to 1.1470 mg/cm³
Traceable Report #: 1540993E
NIST Certificate #: 822/272801-06 822/274081-06
Tested By: 21
Procedure: Mass Dissemination (reference HB 952)
Contractor Req Recall Date: 2 Year
Primary Standard Calibration Date: 11/03/06 **Due:** 11/03/10



★★ Indicates replaced magnetic weight

Description of Weights: 3kg-200g Satin Finish Kit, Class "2", S/N F940 & 300g Satin Finish Weight, Class "2", S/N K303
 Although there are two NIST numbers, one or both may apply.

Nominal Value	Id.	Conventional Mass Corr.		Unc. ¹ K=2 (mg)	Tot. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY		Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)				MM-DD-YY	MM-DD-YY	
200 g		-0.470	0.464	★★ 0.041	1.0	699Q	L595Q	08-10-09/11-10-09	7.84
200 g		-0.690	0.447	★★ 0.041	1.0	699Q	L595Q	08-10-09/11-10-09	7.84
300 g	K303	0.600	0.677	★★ 0.058	1.5	699Q	L595Q	08-10-09/11-10-09	7.84
1 kg		-1.79	-1.79	0.11	5.0	699Q	L595Q	08-10-09/11-10-09	7.84
2 kg		-7.98	-7.98	0.41	10	124Q	L595Q	08-10-09/11-10-09	7.84
3 kg		-13.85	-13.85	0.65	15	124Q	L595Q	08-10-09/11-10-09	7.84

¹ Uncertainties apply to As Left values only

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.



230 West Coleman Street
 Rice Lake, WI 54988 • USA
 TEL: 715-234-9171 • FAX: 715-234-6967

Dated 18 AUG 2009

Dan Demers
 Dan Demers Metrologist

