

**Contractor:** G T Michelli  
 130 Brookhollow Esplanade  
 Harahan, LA 70123 5102  
  
**Purchase Order #:** 11017213  
**Client:** G T Michelli  
**Address:** 130 Brookhollow Esplanade  
  
**City & State:** Harahan, LA 70123 5102  
**Date Received:** 10 AUG 2009  
**Date Calibrated:** 14 AUG 2009  
**Temperature Range:** 21.37 °C  
**Pressure Range:** 731.1 to 731.2 mmHg  
**Relative Humidity Range:** 48 to 49 %  
**Air Density:** 1.1474 to 1.1476 mg/cm<sup>3</sup>  
**Traceable Report #:** 1543646  
**NIST Certificate #:** 822/272801-06 822/274081-06  
**Tested By:** 19  
**Procedure:** Modified Substitution (WI05-0023)  
**Contractor Req Recall Date:** 2 Year  
**Primary Standard Calibration Date:** 11/03/06 **Due:** 11/03/10  
**Description of Weights:** 2-8oz, 2-1lb, 2-2lb Satin Finish Kit, Class "F", S/N 2NUY  
 Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tot. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm <sup>3</sup> )
		As Found (mg)	As Left (mg)					
8 oz	02NH	18.2	18.2	5.1	45	859Q	D564Q 04-29-09/10-29-09	7.84
8 oz	02NI	18.2	18.2	5.1	45	859Q	D564Q 04-29-09/10-29-09	7.84
1 lb	02NT	24.3	24.3	5.7	70	859Q	D564Q 04-29-09/10-29-09	7.84
1 lb	02NU	24.1	24.1	5.7	70	859Q	D564Q 04-29-09/10-29-09	7.84
2 lb	0205	31.0	31.0	5.3	91	859Q	D564Q 04-29-09/10-29-09	7.84
2 lb	0206	31.5	31.5	5.3	91	859Q	D564Q 04-29-09/10-29-09	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

**RICE LAKE**  
 WEIGHING SYSTEMS  
 230 West Coleman Street  
 Rice Lake, WI 54868 • USA  
 TEL: 715-234-9171 • FAX: 715-234-6967



Dated 14 AUG 2009

Debbie Schieffer Lab Supervisor