

Contractor: G T Michelli
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 11017214
Client: G T Michelli
Address: 130 Brookhollow Esplanade

City & State: Harahan, LA 70123 5102
Date Received: 10 AUG 2009
Date Calibrated: 14 AUG 2009
Temperature Range: 21.37 to 21.39 °C
Pressure Range: 731.1 to 731.2 mmHg
Relative Humidity Range: 48 to 49 %
Air Density: 1.1474 to 1.1476 mg/cm³
Traceable Report #: 1543648
NIST Certificate #: 822/272801-06 822/274081-06
Tested By: 19
Procedure: Modified Substitution (WI05-0023)
Contractor Req Recall Date: 2 Year
Primary Standard Calibration Date: 11/03/06 **Due:** 11/03/10
Description of Weights: 2-8oz, 2-1lb, 2-2lb Satin Finish Kit, Class "F", Set S/N KIT A



Although there are two NIST numbers, one or both may apply.

Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
8 oz	02NF	19.2	19.2	5.1	45	859Q	D564Q 04-29-09/10-29-09	7.84
8 oz	02NG	17.2	17.2	5.1	45	859Q	D564Q 04-29-09/10-29-09	7.84
1 lb	02NR	22.6	22.6	5.7	70	859Q	D564Q 04-29-09/10-29-09	7.84
1 lb	02NS	22.5	22.5	5.7	70	859Q	D564Q 04-29-09/10-29-09	7.84
2 lb	0203	31.0	31.0	5.3	91	859Q	D564Q 04-29-09/10-29-09	7.84
2 lb	0204	33.5	33.5	5.3	91	859Q	D564Q 04-29-09/10-29-09	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By: _____ Check with your local state agency for certification of compliance on legal-for-trade items.



230 West Coleman Street
 Rice Lake, WI 54868 • USA
 TEL: 715-234-9171 • FAX: 715-234-6967

Dated 14 AUG 2009

Debbie Schieffer
 Debbie Schieffer Lab Supervisor

