

Contractor: G T Michelli
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 11017235
Client: G T Michelli
Address: 130 Brookhollow Esplanade

City & State: Harahan, LA 70123 5102
Date Received: 20 AUG 2009
Date Calibrated: 24 AUG 2009
Temperature Range: 20.85 to 20.89 °C
Pressure Range: 733.0 mmHg
Relative Humidity Range: 50 to 51 %
Air Density: 1.1523 to 1.1525 mg/cm³
Traceable Report #: 1547170
NIST Certificate #: 822/272801-06 822/274081-06
Tested By: 19
Procedure: Modified Substitution (WI05-0023)
Contractor Req Recall Date: 2 Year
Primary Standard Calibration Date: 11/03/06 **Due:** 11/03/10
Description of Weights: 8oz - 2lb, Satin Finish Weights, Class F



Although there are two NIST numbers, one or both may apply.

Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used	Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)						
8 oz	02NL	19.2	19.2	4.8	45	859Q	D564Q	04-29-09/10-29-09	7.84
8 oz	02NM	18.2	18.2	4.8	45	859Q	D564Q	04-29-09/10-29-09	7.84
1 lb	02NX	18.5	18.5	6.0	70	859Q	D564Q	04-29-09/10-29-09	7.84
1 lb	02NY	30.2	30.2	6.0	70	859Q	D564Q	04-29-09/10-29-09	7.84
2 lb	0209	32.5	32.5	6.1	91	859Q	D564Q	04-29-09/10-29-09	7.84
2 lb	020A	36.8	36.8	6.1	91	859Q	D564Q	04-29-09/10-29-09	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

RICE LAKE
WEIGHING SYSTEMS



230 West Coleman Street
 Rice Lake, WI 54868 • USA
 TEL: 715-234-9171 • FAX: 715-234-6967

Dated 24 AUG 2009

Dan Demers
 Dan Demers Metrologist