

RLWS *Mass Value* CERTIFICATE

Contractor:

G T Michelli
130 Brookhollow Esplanade
Harahan, LA 70123 5102

Purchase Order #:

11017279

Client:

G T Michelli

Address:

130 Brookhollow Esplanade

City & State:

Harahan, LA 70123 5102

Date Received:

04 SEP 2009

Date Calibrated:

09 SEP 2009

Temperature Range:

21.54 to 21.81 °C

Pressure Range:

735.4 to 735.6 mmHg

Relative Humidity Range:

50 %

Air Density:

1.1525 to 1.1537 mg/cm³

Traceable Report #:

1551942

NIST Certificate #:

822/272801-06 822/274081-06

Tested By:

12

Procedure:

Modified Substitution (WI05-0023)

Contractor Req Recall Date:

2 Year

Primary Standard Calibration Date:

11/03/06

Due: 11/03/10

Description of Weights:

500g-5kg Satin Finish Kit, Class "2", S/N 2NUZ

Although there are two NIST numbers, one or both may apply.



Indicates As Found >= Tol

Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
500 g	J874	-1.225	-1.225	0.079	2.5	619Q	K594Q 08-25-09/11-25-09	7.84
500 g	J871	-0.403	-0.403	0.079	2.5	619Q	K594Q 08-25-09/11-25-09	7.84
1 kg	N780	-5.04	1.90	<input checked="" type="checkbox"/> 0.13	5.0	619Q	K594Q 08-25-09/11-25-09	7.84
1 kg	N779	-3.00	-3.00	0.13	5.0	619Q	K594Q 08-25-09/11-25-09	7.84
1 kg	J869	-4.38	-4.38	0.13	5.0	619Q	K594Q 08-25-09/11-25-09	7.84
1 kg		-1.94	-1.94	0.13	5.0	619Q	K594Q 08-25-09/11-25-09	7.84
5 kg	N777	4.36	4.36	0.93	25	975Q	K594Q 08-25-09/11-25-09	7.84
5 kg	N776	4.07	4.07	0.93	25	975Q	K594Q 08-25-09/11-25-09	7.84
5 kg	N778	2.29	2.29	0.93	25	975Q	K594Q 08-25-09/11-25-09	7.84

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:



230 West Coleman Street • Rice Lake, WI 54868 • USA
TEL: 715-234-9171 • FAX: 715-234-6967

Page 1 of 1 Page

Dated 09 SEP 2009

Dan Demers
Dan Demers Metrologist

This report is not to be used to claim product endorsement by Rice Lake Weighing Systems or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory. The above uncertainty does not include a component for magnetic property or handling and use.

PN 38914 6/07