

Contractor:

G T Michelli
130 Brookhollow Esplanade
Harahan, LA 70123 5102

Purchase Order #:

11017274

Client:

G T Michelli

Address:

130 Brookhollow Esplanade

City & State:

Harahan, LA 70123 5102

Date Received:

04 SEP 2009

Date Calibrated:

08 SEP 2009

Temperature Range:

21.21 to 21.70 °C

Pressure Range:

734.2 to 734.3 mmHg

Relative Humidity Range:

50 %

Air Density:

1.1509 to 1.1530 mg/cm³

Traceable Report #:

1551944

NIST Certificate #:

822/272801-06 822/274081-06

Tested By:

12

Procedure:

Modified Substitution (WI05-0023)

Contractor Req Recall Date:

2 Year

Primary Standard Calibration Date:

11/03/06

Due: 11/03/10

Description of Weights:

100g-2kg Satin Finish Kit, Class "2", S/N 020D

Although there are two NIST numbers, one or both may apply.



| Nominal Value | Id. | Conventional Mass Corr. | | Unc. K=2 (mg) | Tol. (mg) | Balance Used | Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY | Assumed Density (g/cm³) |
|---------------|------|-------------------------|--------------|---------------|-----------|--------------|--|-------------------------|
| | | As Found (mg) | As Left (mg) | | | | | |
| 100 g | | -0.325 | -0.325 | 0.038 | 0.50 | 1183Q | K594Q 08-25-09/11-25-09 | 7.84 |
| 200 g | | -0.537 | -0.537 | 0.059 | 1.0 | 619Q | K594Q 08-25-09/11-25-09 | 7.84 |
| 200 g | | -0.305 | -0.305 | 0.059 | 1.0 | 619Q | K594Q 08-25-09/11-25-09 | 7.84 |
| 500 g | | -0.027 | -0.027 | 0.078 | 2.5 | 619Q | K594Q 08-25-09/11-25-09 | 7.84 |
| 500 g | | -0.844 | -0.844 | 0.078 | 2.5 | 619Q | K594Q 08-25-09/11-25-09 | 7.84 |
| 1 kg | N781 | -4.60 | -4.60 | 0.13 | 5.0 | 619Q | K594Q 08-25-09/11-25-09 | 7.84 |
| 1 kg | | -1.83 | -1.83 | 0.13 | 5.0 | 619Q | K594Q 08-25-09/11-25-09 | 7.84 |
| 2 kg | | -4.26 | -4.26 | 0.62 | 10 | 975Q | K594Q 08-25-09/11-25-09 | 7.84 |

Check with your local state agency for certification of compliance on legal-for-trade items.

Prepared By:



Dated 08 SEP 2009

Dan Demers
Dan Demers Metrologist

230 West Coleman Street • Rice Lake, WI 54868 • USA
TEL: 715-234-9171 • FAX: 715-234-6967

This report is not to be used to claim product endorsement by Rice Lake Weighing Systems or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory. The above uncertainty does not include a component for magnetic property or handling and use.