


Contractor: G T Michelli
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 11017273
Client: G T Michelli
Address: 130 Brookhollow Esplanade

City & State: Harahan, LA 70123 5102
Date Received: 04 SEP 2009
Date Calibrated: 09 SEP 2009
Temperature Range: 21.02 to 21.15 °C
Pressure Range: 735.3 mmHg
Relative Humidity Range: 50 to 51 %
Air Density: 1.1549 to 1.1554 mg/cm³
Traceable Report #: 1551945
NIST Certificate #: 822/272801-06 822/274081-06
Tested By: 17
Procedure: Modified Substitution (WI05-0023)
Contractor Req Recall Date: 2 Year
Primary Standard Calibration Date: 11/03/06 **Due:** 11/03/10
Description of Weights: 2-8oz, 2-1lb, 2-2lb Satin Finish Weights, Class "F"
 Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tot. (mg)	Balance Used	Standard Set Used		Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)				Calibrated/due	MM-DD-YY/MM-DD-YY	
8 oz	02NJ	16.4	16.4	5.2	45	859Q	D564Q	04-29-09/10-29-09	7.84
8 oz	02NK	17.8	17.8	5.2	45	859Q	D564Q	04-29-09/10-29-09	7.84
1 lb	02NV	31.0	31.0	5.5	70	859Q	D564Q	04-29-09/10-29-09	7.84
1 lb	02NW	25.0	25.0	5.5	70	859Q	D564Q	04-29-09/10-29-09	7.84
2 lb	0207	28.2	28.2	6.0	91	859Q	D564Q	04-29-09/10-29-09	7.84
2 lb	0208	38.6	38.6	6.0	91	859Q	D564Q	04-29-09/10-29-09	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

RICE LAKE
 WEIGHING SYSTEMS
 230 West Coleman Street
 Rice Lake, WI 54868 • USA
 TEL: 715-234-9171 • FAX: 715-234-6967



Dated 09 SEP 2009

 Dan Demers Metrologist

