

Contractor: G T Michelli
130 Brookhollow Esplanade
Harahan, LA 70123 5102

Purchase Order #: 11017613
Client: G T Michelli
Address: 130 Brookhollow Esplanade

City & State: Harahan, LA 70123 5102

Date Received: 13 JAN 2010

Date Calibrated: 18 JAN 2010

Temperature Range: 20.38 to 20.43 °C

Pressure Range: 729.6 to 729.7 mmHg

Relative Humidity Range: 50 to 51 %

Air Density: 1.1489 to 1.1492 mg/cm³

Traceable Report #: 1586943

NIST Certificate #: 822/274081-06, 822/278785-10

Tested By: 19

Procedure: Modified Substitution (WI05-0023)

Contractor Req Recall Date: 18 JAN 2012

Primary Standard Calibration Date: 11/03/06, 10/21/09 **Due:** 11/03/10, 10/21/13

Description of Weights: 6-5lb Satin Finish Weights, Class "F"

Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used Calibrated/du	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
5 lb	0V0H	58.6	58.6	7.3	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V0I	68.3	68.3	7.3	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V0J	78.0	78.0	7.3	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V0K	84.7	84.7	7.3	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V0L	75.4	75.4	7.3	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V0M	81.5	81.5	7.8	227	859Q	D564Q 10-27-09/04-27-10	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

RICE LAKE
WEIGHING SYSTEMS



230 West Coleman Street
Rice Lake, WI 54868 • USA
TEL: 715-234-9171 • FAX: 715-234-6967

Dated 18 JAN 2010

Dan Demers
Dan Demers Metrologist

