

Contractor: G T Michelli
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: 11017612
Client: G T Michelli
Address: 130 Brookhollow Esplanade

City & State: Harahan, LA 70123 5102
Date Received: 13 JAN 2010
Date Calibrated: 18 JAN 2010
Temperature Range: 20.43 to 20.48 °C
Pressure Range: 729.4 mmHg
Relative Humidity Range: 51 %
Air Density: 1.1483 to 1.1486 mg/cm³
Traceable Report #: 1586944
NIST Certificate #: 822/274081-06, 822/278785-10
Tested By: 19
Procedure: Modified Substitution (WI05-0023)
Contractor Req Recall Date: 18 JAN 2012
Primary Standard Calibration Date: 11/03/06, 10/21/09 **Due:** 11/03/10, 10/21/13
Description of Weights: 6-5lb Satin Finish Weights, Class "F"
 Although there are two NIST numbers, one or both may apply.



| Nominal Value | Id. | Conventional Mass Corr. | | Unc. K=2 (mg) | Tol. (mg) | Balance Used | Standard Set Used | | Assumed Density (g/cm³) |
|---------------|------|-------------------------|--------------|---------------|-----------|--------------|-------------------|-------------------|-------------------------|
| | | As Found (mg) | As Left (mg) | | | | Calibrated/due | MM-DD-YY/MM-DD-YY | |
| 5 lb | 0UZT | 67.4 | 67.4 | 7.8 | 227 | 859Q | D564Q | 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0UZU | 67.6 | 67.6 | 7.8 | 227 | 859Q | D564Q | 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0UZV | 80.7 | 80.7 | 7.8 | 227 | 859Q | D564Q | 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0UZW | 50.8 | 50.8 | 7.8 | 227 | 859Q | D564Q | 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0UZX | 84.0 | 84.0 | 7.8 | 227 | 859Q | D564Q | 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0UZY | 57.8 | 57.8 | 7.7 | 227 | 859Q | D564Q | 10-27-09/04-27-10 | 7.84 |

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.



230 West Coleman Street
 Rice Lake, WI 54868 • USA
 TEL: 715-234-9171 • FAX: 715-234-6967

Dated 18 JAN 2010

Dan Demers
 Dan Demers Metrologist

