

Contractor: G T Michelli
 130 Brookhollow Esplanade
 Harahan, LA 70123 5102

Purchase Order #: .11017648
Client: G T Michelli
Address: 130 Brookhollow Esplanade

City & State: Harahan, LA 70123 5102
Date Received: 27 JAN 2010
Date Calibrated: 01 FEB 2010
Temperature Range: 20.56 to 20.59 °C
Pressure Range: 735.4 mmHg
Relative Humidity Range: 51 %
Air Density: 1.1573 to 1.1574 mg/cm³
Traceable Report #: 1591491
NIST Certificate #: 822/274081-06, 822/278785-10
Tested By: 19
Procedure: Modified Substitution (WI05-0023)
Contractor Req Recall Date: 02-01-2012
Primary Standard Calibration Date: 11/03/06, 10/21/09 **Due:** 11/03/10, 10/21/13
Description of Weights: 6-5lb Satin Finish Weights, Class "F"
 Although there are two NIST numbers, one or both may apply.



| Nominal Value | Id. | Conventional Mass Corr. | | Unc. K=2 (mg) | Tol. (mg) | Balance Used | Standard Set Used Calibrated/due MM-DD-YY/MM-DD-YY | Assumed Density (g/cm ³) |
|---------------|------|-------------------------|--------------|---------------|-----------|--------------|--|--------------------------------------|
| | | As Found (mg) | As Left (mg) | | | | | |
| 5 lb | 0V0B | 53.3 | 53.3 | 8.6 | 227 | 859Q | D564Q 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0V0C | 66.8 | 66.8 | 8.6 | 227 | 859Q | D564Q 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0V0D | 75.4 | 75.4 | 8.6 | 227 | 859Q | D564Q 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0V0E | 39.0 | 39.0 | 8.6 | 227 | 859Q | D564Q 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0V0F | 26.6 | 26.6 | 8.6 | 227 | 859Q | D564Q 10-27-09/04-27-10 | 7.84 |
| 5 lb | 0V0G | 53.8 | 53.8 | 8.2 | 227 | 859Q | D564Q 10-27-09/04-27-10 | 7.84 |

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

RICE LAKE
WEIGHING SYSTEMS

230 West Coleman Street, Rice Lake, WI 54868 • USA
 TEL: 715-234-9171 • FAX: 715-234-6967 • www.ricolake.com
 An ISO 9001 Registered Company

Dated 01 FEB 2010

Dan Demers
 Dan Demers Metrologist