

Contractor:

G T Michelli
130 Brookhollow Esplanade
Harahan, LA 70123 5102

Purchase Order #:

11017696

Client:

G T Michelli

Address:

130 Brookhollow Esplanade

City & State:

Harahan, LA 70123 5102

Date Received:

09 FEB 2010

Date Calibrated:

11 FEB 2010

Temperature Range:

20.53 to 20.59 °C

Pressure Range:

732.4 mmHg

Relative Humidity Range:

50 %

Air Density:

1.1526 to 1.1529 mg/cm³

Traceable Report #:

1595472

NIST Certificate #:

822/274081-06, 822/278785-10

Tested By:

19

Procedure:

Modified Substitution (WI05-0023)

Contractor Req Recall Date:

2 Years


Primary Standard Calibration Date:

11/03/06, 10/21/09 **Due:** 11/03/10, 10/21/13

Description of Weights:

6-5lb, Satin Finished Stainless Steel, Cylindrical Weights, Class F

Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used Calibrated/du	Assumed Density (g/cm ³)
		As Found (mg)	As Left (mg)					
5 lb	0V05	67.7	67.7	7.8	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V06	89.0	89.0	7.8	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V07	80.3	80.3	7.8	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V08	63.6	63.6	7.8	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V09	59.8	59.8	7.8	227	859Q	D564Q 10-27-09/04-27-10	7.84
5 lb	0V0A	38.8	38.8	7.3	227	859Q	D564Q 10-27-09/04-27-10	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

RICE LAKE
WEIGHING SYSTEMS

230 West Coleman Street, Rice Lake, WI 54868 • USA
TEL: 715-234-9171 • FAX: 715-234-6967 • www.ricelake.com
An ISO 9001 Registered Company

Page 1 of 1 Page

Dated 11 FEB 2010

Dan Demers
Dan Demers Metrologist

NVLAP
NVLAP Lab Code 105001-0

This report is not to be used to claim product endorsement by Rice Lake Weighing Systems, NVLAP or any agency of the U.S. Government. This document shall not be reproduced, except in full, without the written approval of Rice Lake Weighing Systems' Metrology Laboratory. For ASTM Class 4, 5, 6, 7 and NIST F the noted uncertainty does not include a component of magnetic properties or handling and use.

PN 64784 12/09