

Contractor: G T Michelli (Har)
130 Brookhollow Esplanade
Harahan, LA 70123 5102

Purchase Order #: 11017902
Client: G T Michelli (Har)
Address: 130 Brookhollow Esplanade

City & State: Harahan, LA 70123-5102
Date Received: 26 APR 2010
Date Calibrated: 29 APR 2010
Temperature Range: 20.40 to 20.56 °C
Pressure Range: 718.6 to 718.7 mmHg
Relative Humidity Range: 50 to 51 %
Air Density: 1.1309 to 1.1317 mg/cm³
Traceable Report #: 1619950
NIST Certificate #: 822/274081-06, 822/278785-10
Tested By: 17
Procedure: Modified Substitution (WI05-0023)
Contractor Req Recall Date: 2 Years
Primary Standard Calibration Date: 11/03/06, 10/21/09 **Due:** 11/03/10, 10/21/13
Description of Weights: 6-5lb Satin Finish Weights, NIST Class F

Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used		Assumed Density (g/cm³)
		As Found (mg)	As Left (mg)				Calibrated/due	MM-DD-YY/MM-DD-YY	
5 lb	0UZZ	76.3	76.3	7.2	227	859Q	D564Q	04-27-10/10-27-10	7.84
5 lb	0V00	58.6	58.6	7.1	227	859Q	D564Q	04-27-10/10-27-10	7.84
5 lb	0V01	94.3	94.3	7.2	227	859Q	D564Q	04-27-10/10-27-10	7.84
5 lb	0V02	76.8	76.8	7.2	227	859Q	D564Q	04-27-10/10-27-10	7.84
5 lb	0V03	71.0	71.0	7.1	227	859Q	D564Q	04-27-10/10-27-10	7.84
5 lb	0V04	67.4	67.4	7.1	227	859Q	D564Q	04-27-10/10-27-10	7.84

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

RICE LAKE
WEIGHING SYSTEMS

230 West Coleman Street, Rice Lake, WI 54868 • USA
TEL: 715-234-9171 • FAX: 715-234-6967 • www.ricelake.com
An ISO 9001 Registered Company

Page 1 of 1 Page

Dated 29 APR 2010

Dan Demers
Dan Demers Metrologist