

**Contractor:**

**G T Michelli Co Inc**  
130 Brookhollow Esplanade  
Harahan, LA 70123 5102

**Purchase Order #:**

11018165

**Client:**

**G T Michelli CO Inc**

**Address:**

130 Brookhollow Esplanade

**City & State:**

Harahan, LA 70123-5102

**Date Received:**

11 AUG 2010

**Date Calibrated:**

13 AUG 2010

**Temperature Range:**

21.59 °C

**Pressure Range:**

724.3 mmHg

**Relative Humidity Range:**

51 %

**Air Density:**

1.1354 mg/cm<sup>3</sup>

**Traceable Report #:**

1654631

**NIST Certificate #:**

822/274081-06, 822/278785-10

**Tested By:**

17

**Procedure:**

Mass Dissemination (reference HB 952)

**Contractor Req Recall Date:**

2 Years

**Primary Standard Calibration Date:**

11/03/06, 10/21/09 Due: 11/03/10, 10/21/13

**Description of Weights:**

2-500mg Stainless Steel Leaf Weights, OIML Class E2

Although there are two NIST numbers, one or both may apply.



Nominal Value	Id.	Conventional Mass Corr.		Unc. <sup>1</sup> K=2 (mg)	Tol. (mg)	Balance Used	Standard Set Used Calibrated/du MM-DD-YY/MM-DD-YY	Assumed Density (g/cm <sup>3</sup> )
		As Found (mg)	As Left (mg)					
500 mg	.	-0.0053	-0.0053	0.0014	0.025	650Q	L595Q 08-05-10/11-05-10	7.95
500 mg	..	-0.0028	-0.0028	0.0014	0.025	650Q	L595Q 08-05-10/11-05-10	7.95

<sup>1</sup> Uncertainties apply to As Left values only

This report contains data not covered by the NVLAP Accreditation if the box is checked.

Prepared By:

Check with your local state agency for certification of compliance on legal-for-trade items.

**RICE LAKE**  
WEIGHING SYSTEMS

230 West Coleman Street, Rice Lake, WI 54868 • USA  
TEL: 715-234-9171 • FAX: 715-234-6967 • [www.ricelake.com](http://www.ricelake.com)  
An ISO 9001 Registered Company

Page 1 of 1 Page

Dated 13 AUG 2010

*Dan Demers*  
Dan Demers Metrologist

